

<b>DATE:</b>	2022 January 14
<b>TO:</b>	Grey Nuns Community Hospital – Physicians, Nurses, Pharmacists, and Managers
<b>FROM:</b>	Alberta Precision Laboratories
<b>RE:</b>	<b>New Coagulation Analyzer Implementation at Grey Nuns Community Hospital Lab</b>

---

## PLEASE POST OR DISTRIBUTE AS WIDELY AS APPROPRIATE

---

### Key Message

- New ACL TOP coagulation analyzers are going live **January 18, 2022** at **Grey Nuns Community Hospital**.
- These new analyzers are being implemented in a staggered fashion across the entire Edmonton Zone over the span of a few weeks. As this platform is already in use everywhere else in Alberta, implementation also allows for provincial standardization of coagulation testing.

### Background

- The coagulation lab at **Grey Nuns Community Hospital** already performs the following tests: **PTT, INR, Fibrinogen, D-Dimer, Anti-Xa Heparin Level and Thrombin Time**
- The new coagulation analyzers will continue to perform all the same tests. However, the new coagulation analyzers, from Instrumentation Laboratories (aka Werfen), utilize a different clot detection method (optical) than the retiring Stago brand of analyzers (mechanical). Consequently, significant changes to how results are reported will occur.
- Additional changes will also be occurring as a result of provincial standardization.

### How this will impact you

#### 1) Interfering Substances:

- Thresholds for interfering substances will change (see **Appendix Table 1**). Comments will continue to be appended if results may be affected by interfering substances.
- Thrombin Time is particularly sensitive to interfering substances and the test will be cancelled if there is visible lipemia or hemolysis.

#### 2) Reference Ranges:

- Please refer to **Appendix Table 2** for summary of changes.

#### 3) Heparin Monitoring:

- Continue to use Anti-Xa Heparin Nomogram for heparin monitoring per previous bulletin (PTT is not validated for heparin monitoring).

#### 4) Turnaround Times:

- On-instrument analysis time may be prolonged in the setting of significantly abnormal INR, PTT and Fibrinogens. This may lead to extended turnaround time for critical results.
- The *maximum* differences from our previous analyzers have been estimated as follows:
  - INR (when >5): +3 minutes.
  - PTT (when >120 s): +7 minutes
  - Fibrinogen (when <1.5 g/L): +4 minutes

#### 5) D-Dimer Assay Changes

- The new D-Dimer assay (HemosIL D-Dimer HS500) maintains the same clinical cut-off for exclusion of venous thromboembolism as the previous assay (Stago STA-Liatest D-Di) of 0.50 mg/L FEU.



- Otherwise, however, the assays are by no means equivalent. The incoming HemosIL HS-500 assay reactivity is progressively higher than the retiring Stago assay for values >0.50 mg/L.
- This stems from a lack of an international standard for D-Dimer calibration.
- For example, for a sample run in parallel on both analyzers:
  - Stago STA-Liatest D-Di result: 5.0 mg/L
  - HemosIL HS-500 result: 7.5 mg/L
- Thus, we advise AGAINST adjusting the cut-off for venous thromboembolism exclusion based on patient age.
- **Note:** Most widely available age-adjustment calculators online do not account for differences between assay types/manufacturers.

#### **6) Critical Results - Preliminary Reporting and Reflex Testing:**

- Provincial standardization initiatives coincide with the new analyzer implementation and necessitate changes to the process for laboratory handling of critical results.
- Previously, critical PTTs, INRs, and fibrinogens were *Preliminarily* reported while the laboratory confirmed their validity, and then were *Final Verified*.
- Going forward, only critical Fibrinogen assay results will be *Preliminarily* reported. The lab will then investigate for possible interferences and if none are found, the Fibrinogen result will be *Final Verified*. INR and PTT results will only be available at the *Final Verified* result stage.
- Under some circumstances, a critical INR or PTT result may cause a reflex Fibrinogen assay to be run. In these cases, a comment will be appended to the INR/PTT result indicating testing was added-on.

#### **7) Critical Results – Contacting Collectors:**

- The new critical handling process also acknowledges the risk of specimen collection issues causing expectedly grossly abnormal (critical) results.
- Laboratory technical staff may attempt to contact the specimen collector (nursing, allied health, lab phlebotomist, etc.) to inquire regarding the collection if the results were not previously critical.
- Comments may be appended to the results if the lab suspects or cannot determine if there was specimen interference/compromise.

#### **Action Required**

- Be aware of limitations and changes associated with incoming coagulation analyzers relevant to your practice.

#### **Effective**

- January 18, 2022

#### **Questions/Concerns**

- Dr. Dean Tung, Grey Nuns Community Hospital Lab Site Lead, [dean.tung@albertaprecisionlabs.ca](mailto:dean.tung@albertaprecisionlabs.ca)
- Dr. E. Turley, Coagulation Lead - Edmonton, [Elona.Turley@albertaprecisionlabs.ca](mailto:Elona.Turley@albertaprecisionlabs.ca)

#### **Approved by**

- Dr. A. Szkotak, Hematopathology Chief, North Sector, [Artur.Szkotak@albertaprecisionlabs.ca](mailto:Artur.Szkotak@albertaprecisionlabs.ca)



## APPENDIX

**Table 1.:** Tests on the ACL TOP instruments are accurate up to (and including) the tabulated levels of the interfering substance

Test	Interfering Substance			
	Triglycerides (mmol/L)	Bilirubin (umol/L)	Free Hemoglobin (mg/L hemolysis)	Heparin (U/mL)
INR	10	510	5000	1.0
PTT	10	425	5000	N/A – sensitive to heparin
Fibrinogen	8	350	3750	1.0
D-Dimer	15	305	5000	10
Thrombin Time	N/A – sensitive to lipemia	410	N/A – sensitive to hemolysis	N/A – sensitive to heparin
Anti-Xa Heparin	10	480	5000	N/A – sensitive
Antithrombin 3	26	680	5000	4.0

**Table 2:** Changes to Reference Ranges (Applicable to RAH, MCH, & GNH)

Test	New Reference Range
INR	Unchanged
PTT	Unchanged
Fibrinogen	2.0 – 4.0 g/L
D-Dimer	Unchanged
Thrombin Time	10.3 – 16.6 seconds
Anti-Xa Heparin	Unchanged
Antithrombin 3	>=0.80