

Connect Care Cleaning and Disinfecting of Mobile Electronic Devices

Note: For more information, refer to the IPC Best Practice Recommendation for [Bedside Computers and Electronic Devices](#) and the [Steps for Disinfecting Telehealth Equipment and Peripheral Devices Poster](#). If you have any questions or comments contact IPC at ipcsurvstdadmin@ahs.ca.




This work is licensed under a [Creative Commons Attribution-Non-commercial-Share Alike 4.0 International license](#). The licence does not apply to AHS trademarks, logos or content for which Alberta Health Services is not the copyright owner.

Disclaimer: This material is intended for general information only and is provided on an "as is", "where is" basis. Although reasonable efforts were made to confirm the accuracy of the information, Alberta Health Services does not make any representation or warranty, express, implied or statutory, as to the accuracy, reliability, completeness, applicability or fitness for a particular purpose of such information. This material is not a substitute for the advice of a qualified health professional. Alberta Health Services expressly disclaims all liability for the use of these materials, and for any claims, actions, demands or suits arising from such use.

Best practice recommendations

Mobile electronic devices may include: workstations on wheels (WOW), rovers, tablets, laptops, pagers, phones or personal digital assistants.

Hand hygiene is the most important factor in preventing transmission of microorganisms



- Disinfect all devices before the first use in your shift
- Disinfect when returning a device to a charging station or storage area
- Disinfect all devices when visibly soiled and as outlined in table below
- Replace electronics when grossly contaminated e.g., saturated with blood or body fluids or cleaning doesn't remove visible soil

Note: High touch means those surfaces that have frequent contact with hands.

Mobile electronic device	Additional cleaning	Disinfectant products
Laptops, scanners (including rovers), signature pads, tablets	Between patients	Follow manufacturer's instructions for use (MIFU). Examples of products that may be indicated in the MIFU: <ul style="list-style-type: none"> • Alcohol swabs and wipes • Combination products such as alcohol/quaternary ammonium e.g., CaviWipes® • Hydrogen peroxide products such as accelerated hydrogen peroxide (AHP) products e.g., Virox RTU®, Accel®, Oxivir®
Workstation on Wheels (WOW)	Between patients if: <ul style="list-style-type: none"> • WOW is used within 2 meters of the patient or • WOW is used with contaminated hands (e.g., during a care process) 	Note: Do not use (AHP) products such as Oxivir TB™ on rovers as they will damage the plastic.

For more information contact ipcsurvstdadmin@ahs.ca
© 2021 Alberta Health Services, IPC

Version	Date (YYYY-MM-DD)
Created	2019-09-25
Updated	2020-05-31
Revised	2021-02-24