

Hydrocollators in Community-based Services

- Always perform hand hygiene before accessing supplies.
- Using a low-level disinfectant, wipe the outside of the hydrocollator daily. Clean the outside of the hydrocollator using the two-step cleaning and disinfection process. Let air dry.
- Clean the inside of the hydrocollator according to manufacturer written instructions, and if visibly soiled.
- Maintain a cleaning and daily water temperature log.